



LI484LED Corridor Light Specification Sheet

IL865
Section C
Rev. 1 - 03/2009



LI484LED Corridor Light

ARCHITECTS' AND ENGINEERS' SPECIFICATION

Corridor Lights shall be TekTone®'s LI484LED or approved equal. Corridor lights shall be sectional in design, and shall include one each amber, green and red flame-retardant translucent polycarbonate plastic lenses, and one white flame-retardant translucent polystyrene plastic lens. All lenses shall be mounted on a flame-retardant (UL® 94V-0) plastic base panel. The lenses shall be individually removable. The lights shall provide indication of patient call priority and staff presence. Wiring connections shall be RJ45 plus power and system common, when used with the Tek-CARE®400 Nurse Call system, and via IDC header/pigtail when used with IR300- or SF380-series devices.

The LI484LED shall be UL® 1069 Listed and cUL® Listed.

FEATURES

The LI484LED 4-LED Corridor Light provides visual indications of calls originating from patient, staff, duty, bath, emergency, presence and code call stations. Patient priority and staff presence are indicated by amber, green, red and white colored lenses. When used with a patient or staff station, the LI484LED indicates all calls originated from that room. When used with a multipurpose or duty station, the LI484LED may be used as a zone light.

- UL® 1069 Listed and cUL® Listed
- RJ45 connections plus power & system common *-or-* header/pigtail
- Snap-in colored lenses
- No bulb replacement needed
- Flame-retardant (UL® 94V-0) plastic
- Excellent visibility
- Supervised power lead

SPECIFICATIONS

Dimensions:	Height:	4.56"	(115 mm)
	Width:	4.56"	(115 mm)
	Depth:	3.3"	(84 mm)
Power:	24-28 VDC, 120 mA		
Finish:	Amber, green and red lenses are flame-retardant translucent polycarbonate plastic. White lens is flame-retardant translucent polystyrene plastic. All are mounted on a flame-retardant (UL® 94V-0) plastic panel.		
Connections:	RJ45 plus power & system common (with Tek-CARE®400), and via IDC header/pigtail (with IR300- or SF380-series devices).		
Housing:	Standard single-gang or dual-gang box and ring.		

REQUIRED COMPONENTS

Tek-CARE®400	Nurse Call System
IR400-, SF480-series	Stations
<i>-or-</i>	
Tek-CARE®NC300II	Nurse Call System
IR300-, SF380-series	Stations

www.tektone.com

277 Industrial Park Road • Franklin, NC 28734 • Phone (828) 524-9967 • Fax (828) 524-9968

Email: tektone@tektone.net • Technical Assistance Email: teksupt@tektone.net

TekTone®'s quality system is registered by UL® to the ISO 9001:2000 standard. (File #A10766.)