

Features

- UL® 1069 Listed
- Simple, reliable operation
- Call-placed indicator
- Call activates light on annunciator panel, corridor light and/or zone light.
- Corridor zone light capability
- Mounted on attractive flame retardant (UL® 94V-0) plastic panel

Description

The SF156B Code Station is used to operate all system code signals (except supervised code blue applications). It may be used independently or connected with other call stations.

Placing a code call is accomplished by pulling the slide-switch located at the center of the panel to the down position. Returning the slide-switch to the up position cancels the call.

Listings



UL® 1069 listed; UL® File Number E71599.



SF156B Code Station

Specifications

Dimensions:	Height:	4.5"	(114 mm)
	Width:	2.75"	(70 mm)
	Depth:	0.75"	(20 mm)
Construction:	Reliable components mounted on an attractive flame retardant (UL® 94V-0) plastic panel.		
Connections:	Plug-in wiring pigtail.		
Housing:	Single gang ring, or single gang ring with double gang box.		

Required Components

Tek-CARE® NC110, NC150 or NC200 Nurse Call System
PK153 Third Priority Control Unit